AMPROBE°

Data Sheet

ACDC-3400 IND CAT IV Industrial True RMS Clamp Meter

This CAT IV rated clamp is ideal for industrial applications and utilities that require an extra level of safety. Includes True RMS sensing for accuracy and dependability. Extra large jaw to accommodate wide diameter wires.

- **■** True RMS
- Measures AC Current up to 1000 ACA, Voltage up to 750 VAC / 1000 VDC, Frequency, Resistance and Capacitance and Duty Cycle
- Peak Hold
- 1000A DC Current
- **■** Relative Zero
- Audible continuity
- Auto and manual ranging
- Auto power off
- Data hold
- Backlight
- Diode Test
- Duty Cycle
- Accommodates conductors up to 2" (51mm) in diameter
- Safety CAT IV 600 V, CAT III 1000 V
- Test leads, battery (installed), Users Manual, and Carrying Case included with the product



No hassle warranty

No waiting.





(note: \$500 MSLP limit)









ACDC-3400 IND CAT IV Industrial True RMS Clamp Meter

Data Sheet

 $\label{eq:Specifications} \textbf{Specifications} \ \, \text{(valid for 23 °C ± 5 °C, for less than 75\% relative humidity)}.$

DC Volts	Ranges		Accuracy			
	660.0 mV, 6.600 V, 66.00 V, 660.0 V, 1000 V ± (0.5% rdg + 2 dgts)					
Input impedance	660 mV: >100 MΩ; 6.6 V	660 mV: >100 MΩ; 6.6 V:10 MΩ; 66 V to 1000 V: 9.1 MΩ				
Overload protection	1000 VDC or 750 VAC rn	1000 VDC or 750 VAC rms				
AC Volts True RMS	Ranges	Frequency	Accuracy			
	660.0 mV	50 to 100 Hz	± (1.5% rdg + 8 dgts)			
	6.600V, 66.00V	50 to 500 Hz	± (1.5% rdg + 8 dgts)			
	660.0V, 750V	50 to 500 Hz	± (1.5% rdg + 8 dgts)			
	Frequency*	50 to 1 kHz	± (0.1% rdg + 5 dgts)			
	* Frequency: 10% to 100	0% of voltage rang	ge			
Peak Hold	Ranges	Frequency	Accuracy			
	66.00V, 660.0V, 750V	50 to 500 Hz	± (3.0% rdg + 300 dgts)			
AC coupled TRMS	5% to 100% of range					
Crest factor	≤3					
Input impedance	660 mV: >100 MΩ; 6.6 V:10 MΩ; 66 V to 750 V: 9.1 MΩ					
Overload protection	1000 VDC or 750 VAC rms					
AC Current True RMS	Ranges	Frequency	Accuracy			
	660.0A	50 to 60 Hz	± (2.0% rdg + 10 dgts)			
	660.0A	61 to 400 Hz	± (3.0% rdg + 10 dgts)			
	1000 A	50 to 60 Hz	± (2.5% rdg + 10 dgts)			
	1000 A	61 to 400 Hz	± (3.5% rdg + 10 dgts)			
	Frequency*	50 to 1kHz	± (0.1% rdg + 5 dgts)			
	* Frequency: 10% to 100	0% of current rang	e			
Peak Hold	Ranges	Frequency	Accuracy			
	660.0A, 1000A	50 to 400 Hz	± (3.0% rdg + 200 dgts)			
AC coupled TRMS	5% to 100% of range		(* * * * * * 5 * * * 5 * * * 5 *			
Crest factor	≤ 3					
Overload protection	1000A AC					
DC Current	Range	Accuracy				
	660.0A	± (2.0% rdg +	5 dats)			
	1000A	± (3.0% rdg +	<u>-</u>			
Overload protection	1000A DC	·	<i>5 ·</i>			
Resistance	Range		Accuracy			
	660.0Ω, 6.600kΩ, 66.00kΩ, 660.0kΩ		± (1.0% rdg + 5 dgts)			
	$6.600 ext{M}\Omega$		± (2.0% rdg + 5 dgts)			
	66.00MΩ		± (3.5% rdg + 5 dgts)			
Open circuit volts	-0.8 Vdc typical , (-1.2 Vdc on 660 Ω range)					
Overload protection	1000 VDC or 750 VAC rn					



ACDC-3400 IND CAT IV Industrial True RMS Clamp Meter

Data Sheet

Specifications (continued)

-p	- /			
Capacitance	Range	Accuracy		
	6.600ηϜ, 66.00ηϜ	±(3.0% rdg + 20 dgts)	·	
	660.0ղF, 6.600µF, 66.00µF,660.0µF	±(3.0% rdg + 10 dgts)		
	6.6 mF	±(5.0% rdg + 10 dgts)		
Overload protection	1000 VDC or 750 VAC rms			
Diode Test				
Test current	1.0 mA (approximate)			
Accuracy	± (1.5% rdg + 5 dgts)			
Open circuit volts	3.2 Vdc typical			
Audible indication	< 0.25 V			
Overload protection	1000 VDC or 750 VAC rms			
Continuity				
Ranges	660.0 Ω			
Audible indication	< 30 Ω			
Response time	500 ms			
Overload protection	1000 VDC or 750 VAC rms			
Frequency (Auto ranging)	Range		Accuracy	
	66.00 Hz, 660.0 Hz, 6.600k Hz, 66.00 kH	Hz, 660.0 kHz, 1.000 MHz	± (0.1% rdg + 3 dgts)	
Sensitivity	10 Hz to 1 MHz: > 3.5 V rms			
Minimum pulse Width	>1 us			
% Duty Cycle				
Range	5.0 % to 95.0 %			
Resolution:	0.1 %			
Minimum Pulse Width	>10 us			
Frequency range	40 Hz to 20 kHz			
Accuracy (5V logic)	± (2% rdg + 10 dgts)			
Overload protection	1000 VDC or 750 VAC rms			



ACDC-3400 IND CAT IV Industrial True RMS Clamp Meter

Data Sheet

Technical Data – General Information

Display	3¾ digit liquid crystal display (LCD) (6600 count) with a 66-segment analog bar-graph			
Polarity	Automatic, positive implied, negative polarity indication			
Over range	(OL) or (-OL) is displayed			
Zero	Automatic			
Low battery indication	" is displayed when the battery voltage drops below the operating level			
Auto power off	Approx. 30 minutes			
Backlight	Backlight auto-off approx. 60 sec.			
Measurement rate	2.8 times per second, nominal			
Analog bar-graph	28 times per second			
Operating environment	0°C to 50°C (32°F to 122°F) at < 70% R.H.			
Storage temperature	-20°C to 60°C (-4°F to 140°F) at < 80% R.H. with battery removed from meter			
Temperature Coefficient	0.1 × (specified accuracy) per °C. (0°C to 18°C, 28°C to 50°C)			
Environmental	2000m (6561.7 Feet), Indoor use			
Jaw opening capability	57 mm (2.0 in) conductor			
Power	Single standard 9-volt battery, NEDA 1604, JIS 006P, IEC 6F22			
Battery life	Typically 75 hours with carbon-zinc; 150 hour with alkaline			
Dimensions	283 x 105 x 50 mm (11.1 x 4 x 2.1 in.)			
Weight	559 gm (1.23 lb.)			
Safety	LVD Meets EN61010-1:2001 and EN61010-2-032:2002, CAT III 1000V, CAT IV 600V, class II and pollution degree 2			
С€ ЕМС	EN 61326-1:2006 This product complies with requirements of the following European Community Directives: 2004/108/EC (Electromagnetic Compatibility) and 2006/95/EC (Low Voltage) as amended by 93/68/EEC (CE Marking). However, electrical noise or intense electromagnetic fields in the vicinity of the equipment may disturb the measurement circuit. Measuring instruments will also respond to unwanted signals that may be present within the measurement circuit. Users should exercise care and take appropriate precautions to avoid misleading results when making measurements in the presence of electronic interference.			



Optional Accessories

TL-1500 Test leads with set of alligator clips



Calle del Ebano #16625 Jardines de Chapultepec Tijuana B.C. Mexico Tel. (664) 681 1130 Fax. (664) 681 1150 Tel. 01800 027-4848

©2009 Amprobe Test Tools. All rights reserved. 03/2009 3465184 Rev A