



FEATURES:

- ✓ Designed for all DMM's with a safety style sheathed banana jack.
- ✓ .080" (2mm) notched pin probe tip with screw on standard 4mm banana plug adapter.
- ✓ Probe handle made of Santoprene[®] rubber and Polypropylene for electrical protection, ruggedness and handling comfort.
- ✓ Probe body and banana plug are molded directly to the wire for superior pull strength.
- ✓ Test leads are silicone, high wire strand count for extreme flexibility and high temperature resistance.
- ✓ Banana adapter and banana plug spring are nickel plated beryllium copper for long insertion life.

MATERIALS:

Probe Body:	Gray Santoprene [®] rubber handle highlighted in black or red Polypropylene
Tip:	.080" (2mm) nickel plated brass with 4-48 threads for banana plug adapter
Wire:	18 AWG Silicone insulation, Length: 48" (1,2m)
Banana Plug: Body:	Brass, Nickel Plated
Banana Plug Lantern Spring :	Beryllium Copper, Nickel Plated

RATINGS:

IEC 1010-2-031, Category III

Operating Voltage: 1000 VRMS/DC

Current: 10 Amperes

UL Rated

ORDERING INFORMATION:

Model 6601, includes one black and one red test lead.

All dimensions are in inches. Tolerances (except noted): .xx = ±.02" (.51 mm), .xxx = ±.005" (.127 mm).

All specifications are to the latest revisions. Specifications are subject to change without notice.

Registered trademarks are the property of their respective companies.

Pomona[®]

The items included with this product are certified to meet IEC 1010-2-031 safety requirements.

Les éléments inclus avec ce produit sont certifiés pour répondre à des exigences de sûreté du IEC 1010-2-031.

Los items incluidos con este producto se certifican para resolver requisitos de seguridad del IEC 1010-2-031.

– Only IEC 1010 type shrouded test plugs should be used with IEC 1010 type jacks. NEVER INSERT A PROBE TIP WITH BANANA SPRING INTO A TEST ACCESSORY JACK THAT WILL BE USED AS A HAND HELD PROBING DEVICE

–Seuls les vérificateurs de tension renforcés de type IEC 1010 devraient être utilisés avec des jacks de type IEC 1010. NE JAMAIS INSERER DE BOUT DE SONDE AVEC RESSORT BANANE DANS UN JACK D'ESSAI D'ACCESSOIRES QUI SERA UTILISE COMME DISPOSITIF DE SONDAGE A MAIN.

–Se deben usar sólo tapones de prueba reforzados del tipo IEC 1010 con receptáculos de tipo IEC 1010. NUNCA INTRODUZCA UNA CLAVIJA CON UN MUELLE TIPO BANANA EN UN RECEPTACULO DE PRUEBA QUE A SU VEZ SE USE COMO APARATO DE PRUEBA MANUAL.



Both Direct and Alternating Current.
Courant Continu et Alternatif.
Corriente Directa y Alterna.



Caution ! Refer to Accompanying Documents.
Attention ! Se Reporter aux Documents joints.
; Peligro! Consulte los Documentos Adjuntos.



Double Insulation or Reinforced Insulation.
Isolation Double ou Isolation Renforcée.
Aislamiento Doble o Aislamiento Reforzado.



Caution ! Risk of Electric Shock.
Attention ! Risque de Choc Électrique.
; Peligro! Riesgo de Sacudida Eléctrica.

CAT

Overvoltage Category I, II, III or IV.
Surtension Catégorie I, II, III ou IV.
Sobretensión de Categoría I, II III o IV.

Pomona Electronics



PO Box 9090, Everett, WA 98206-9090 TEL: 800-490-2361

All dimensions are in inches. Tolerances (except noted): .xx = ±.02" (.51 mm), .xxx = ±.005" (.127 mm).

All specifications are to the latest revisions. Specifications are subject to change without notice.

Registered trademarks are the property of their respective companies.

Distribuido Por:

www.finaltest.com.mx

Calle del Ebano #16625

Tijuana B.C. Mexico

Tel. (664)681-1130

Tel. 01800 027-4848

ventas@final-test.net

FINAL TEST[™]
Venta de Instrumentos de Prueba y Medición